

## PROFICIENCY TEST « RAEMA »



réseau d'analyses et d'échanges en microbiologie des aliments

### SCHEME N° 74 (7th MARCH 2022) GENERAL REPORT

« Any reproduction of the report must be made in its entirety»  
« The Cofrac logo may not be used outside this report»

**V. CARLIER<sup>(1)</sup>, L. ALI-MANDJEE et M. CARLIER**  
ASA - ENVA, 7 avenue du Général de Gaulle, 94704 MAISONS ALFORT CEDEX

<sup>(1)</sup> Coordinator of the proficiency test « RAEMA »

For any claim, you can use the specific  
form available on our Website  
<https://association.asa-spv.fr>

## Table of contents

<b>1- GENERAL DATA .....</b>	<b>3</b>
<b>1-1 PARTICIPATING LABORATORIES.....</b>	<b>3</b>
<b>1-2 DELIVERY TIME OF THE PARCEL.....</b>	<b>3</b>
<b>1-3 INFORMATION ABOUT SAMPLE .....</b>	<b>3</b>
<b>1-3-1 NATURE .....</b>	<b>3</b>
<b>1-3-2 SIZE .....</b>	<b>3</b>
<b>1-3-3 HOMOGENEITY AND STABILITY OF THE CONTAMINATION .....</b>	<b>3</b>
<b>1-3-4 FLORA FOR ENUMERATION / DETECTION .....</b>	<b>3</b>
<b>1-4 EXECUTION OF ANALYSIS .....</b>	<b>4</b>
<b>1-4-1 DELIVERY TIME OF SAMPLES / BEGINNING OF ANALYZES.....</b>	<b>4</b>
<b>1-4-2 PRESERVATION TEMPERATURE OF SAMPLE BEFORE ANALYSIS.....</b>	<b>4</b>
<b>2- EXPLOITATION OF ANALYSIS REPORT .....</b>	<b>4</b>
<b>2-1 PREPARATION OF THE INITIAL SUSPENSION .....</b>	<b>4</b>
<b>2-2 DILUENT USED FOR THE INITIAL SUSPENSION .....</b>	<b>4</b>
<b>2-3 HOMOGENEIZATION TECHNIQUE .....</b>	<b>4</b>
<b>2-4 RESUSCITATION'S CONDITIONS .....</b>	<b>4</b>
<b>2-4-1 DURATION .....</b>	<b>4</b>
<b>2-4-2 TEMPERATURE.....</b>	<b>4</b>
<b>2-5 MICROORGANISMS AT 30°C .....</b>	<b>5</b>
<b>2-6 ENTEROBACTERIACEA.....</b>	<b>6</b>
<b>2-7 TOTAL COLIFORMS .....</b>	<b>7</b>
<b>2-8 THERMOTOLERANT COLIFORMS.....</b>	<b>8</b>
<b>2-9 ESCHERICHIA COLI.....</b>	<b>9</b>
<b>2-10 ANAEROBIC SULFITE-REDUCING BACTERIA .....</b>	<b>10</b>
<b>2-11 CLOSTRIDIUM PERFRINGENS .....</b>	<b>11</b>
<b>2-12 COAGULASE POSITIVE STAPHYLOCOCCI .....</b>	<b>12</b>
<b>2-13 LISTERIA MONOCYTOGENES – ENUMERATION .....</b>	<b>13</b>
<b>2-14 SALMONELLA –DETECTION.....</b>	<b>15</b>
<b>2-15 LISTERIA MONOCYTOGENES –DETECTION .....</b>	<b>17</b>
<b>3- ASSESSMENT OF PERFORMANCE .....</b>	<b>19</b>
<b>3-1 PERFORMANCES IN ENUMERATION.....</b>	<b>19</b>
<b>3-1-1 MICROORGANISMS AT 30°C .....</b>	<b>21</b>
<b>3-1-2 ENTEROBACTERIACEA.....</b>	<b>21</b>
<b>3-1-3 TOTAL COLIFORMS .....</b>	<b>21</b>
<b>3-1-4 THERMOTOLERANT COLIFORMS .....</b>	<b>22</b>
<b>3-1-5 ESCHERICHIA COLI.....</b>	<b>22</b>
<b>3-1-6 ANAEROBIC SULFITE-REDUCING BACTERIA .....</b>	<b>22</b>
<b>3-1-7 CLOSTRIDIUM PERFRINGENS .....</b>	<b>23</b>
<b>3-1-8 COAGULASE POSITIVE STAPHYLOCOCCI .....</b>	<b>23</b>
<b>3-1-9 LISTERIA MONOCYTOGENES – ENUMERATION.....</b>	<b>23</b>
<b>3-2 PERFORMANCES IN DETECTION .....</b>	<b>24</b>
<b>3-2-1 DETECTION – SALMONELLA .....</b>	<b>24</b>
<b>3-2-2 DETECTION - LISTERIA MONOCYTOGENES.....</b>	<b>24</b>
<b>3-3 EVOLUTION OF PERFORMANCE .....</b>	<b>24</b>

## 1-GENERAL DATA

### 1.1. PARTICIPATING LABORATORIES

348 laboratories participated to the 74<sup>th</sup> scheme. The sending was made on Monday 7th March 2022. We received 342 answers (98.3%).

### 1.2. DELIVERY TIME OF THE PARCEL

Reception	J0	J0+1	J0+2	J0+3	J0+4	J0+5	J0+7	J0+8	J0+11
Nb of laboratories	7	214	70	20	14	1	9	5	1

### 1.3. INFORMATIONS ABOUT SAMPLE

#### 1.3.1. NATURE

The sample included :

- one strain of *Enterococcus sp.* at a concentration level of  $10^5$  cfu/g in 5 units ;
- one strain of *Citrobacter sp.* at a concentration level of  $10^3$  cfu/g in 5 units ;
- one strain of *Serratia liquefaciens* at a concentration level of  $10^3$  cfu/g in 5 units ;
- one strain of *Escherichia coli* at a concentration level of  $7.10^2$  cfu/g in 5 units ;
- one strain of *Clostridium perfringens* at a concentration level of  $10^2$  cfu/g in 2 units ;
- one strain of *Staphylococcus aureus* at a concentration level of  $5.10^3$  cfu/g in 5 units ;
- one strain of *Salmonella Anatum* at a concentration level of 25 cfu/g in 3 units ;
- one strain of *Listeria monocytogenes* at a concentration level of  $3.10^3$  cfu/g in 3 units .

#### 1.3.2. SIZE

180 kilogrammes of powder were produced and distributed after contamination in bottles containing 75 grammes at least. Bottles were covered by a label with a 6 digit identification number.

#### 1.3.3. HOMOGENEITY AND STABILITY TEST OF THE CONTAMINATION

Homogeneity and stability of samples are checked during the statistical analysis of participants results. A supplementary check of the contamination's homogeneity was realized on 10 samples for each unit by a double enumeration of aerobic microorganisms at 30°C.

The contamination's stability was also checked by enumeration / detection of all flora on 14, 21 and 28 March 2022. These checks were realized by a subcontractor accredited by Cofrac. Homogeneity and stability of samples have been validated.

#### 1.3.4. FLORA FOR ENUMERATION / DETECTION

Enumeration of the following flora was proposed : microorganisms at 30°C, Enterobacteriaceae, total and thermotolerant coliforms, beta-glucuronidase positive *Escherichia coli*, anaerobic sulfite-reducing bacteria, *Clostridium perfringens*, coagulase positive staphylococci, *Listeria monocytogenes*, as well as detection of *Salmonella* and *Listeria monocytogenes*.

## 1.4. EXECUTION OF ANALYZES

### 1.4.1. DELIVERY TIME OF SAMPLES / BEGINNING OF ANALYZES

342 laboratories (100%) specified it.

Analysis time	J0+1	J0+2	J0+3	J0+4	J0+5	J0+7	J0+8	J0+9	J0+10	J0+11	J0+14	J0+15	J0+18
Nb of laboratories	34	57	37	5	1	128	52	8	4	2	9	4	1

### 1.4.2. PRESERVATION TEMPERATURE OF SAMPLES BEFORE ANALYSIS

339 laboratories (99.1%) specified it. The average temperature is **3.9°C** with a standard deviation of 0.9°C. The given data 17, 20, 20.1, 22, 22.5, 24.1 and 25°C given by 11 laboratories were not taken into account for this calculation.

## 2. EXPLOITATION OF ANALYSIS REPORT

### 2.1. PREPARATION OF THE INITIAL SUSPENSION

For 342 answers (100%) :

217 laboratories (63.4%) prepare the initial suspension with adding diluent to powder.  
124 laboratories (36.3%) prepare the initial suspension with adding powder to diluent.  
1 laboratory (0.3%) used another technique to prepare the initial suspension.

### 2.2. DILUENT USED FOR THE INITIAL SUSPENSION

For 338 answers (98.8%) :

301 laboratories (88.0%) use Buffered Peptone Water (or equivalent) for the initial suspension.  
35 laboratories (10.2%) use Peptone salt for the initial suspension.  
2 laboratories (0.6%) used another diluent for the initial suspension.

### 2.3. HOMOGENEIZATION TECHNIQUE

For 339 answers (99.1%) :

314 laboratories (91.8%) homogenize their sampling with a Stomacher<sup>ND</sup>.  
18 laboratories (5.2%) used a manual homogenization.  
3 laboratories (0.9%) used a Vortex mixer.  
4 laboratories (1.2%) used another technique.

## 2.4. RESUSCITATION'S CONDITIONS

### 2.4.1. DURATION

330 laboratories (96.5%) specified it.

The average duration is **26.9 min** with a standard deviation of 15.2 min. The data 90, 120 and 1440 min given by 4 laboratories were not taken into account for this calculation.

### 2.4.2. TEMPERATURE

330 laboratories (96.5%) specified it.

The average temperature is **21.4°C** with a standard deviation of 3.4°C.

## 2.5. MICROORGANISM AT 30°C

328 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	ISO/NF EN ISO 4833-1	211
	AFNOR 3M-01/1-09/89	52
	NM ISO 4833-1	22
	ISO/NF EN ISO 4833-2	11
	AFNOR BIO-12/35-05/13	9
	XP V08-034	8
	Internal method	6
	Other	8
	+ V08-100 (spiral)	19
<b>Culture medium</b>		
	Plate Count Agar	243
	Petrifilms	52
	Plate Count Agar + Milk	23
	Tempo AC	9
	Other	1
<b>Preparation</b>		
	Home made	112
	Ready to use not pre-poured	143
	Ready to use, plate, film, card	71
<b>Plating method</b>		
	Surface	67
	Pour	246
	Culture medium for card	10
<b>1<sup>st</sup> dilution retained</b>		
	- 1	11
	- 2	16
	- 3	175
	- 4	113
	- 5	3
	1/400	5
	1/4000	1
<b>Incubation temperature</b>		
	30°C	324
	33-35°C	2
	37°C	1
	25°C	1
<b>Incubation duration</b>		
	68-74 h	269
	40-48 h	56
	24-26 h	3

## 2.6. ENTEROBACTERIACEA

288 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	NF V08-054	109
	→ NM 08.0.109 <sup>(1)</sup>	18
	ISO/NF EN ISO 21528-2	77
	AFNOR 3M-01/6-09/97	46
	AFNOR BIO-12/21-12/06	10
	AFNOR AES-10/07-01/08	9
	NM ISO 21528-2	8
	AFNOR BRD-07/24-11/13	7
	Internal method	2
	Other	2
<b>Culture medium</b>		
	VRBG	211
	Petrifilms	48
	Tempo EB	10
	Rebecca	10
	Rapid'Enterobacteriaceae	8
	Other	1
<b>Preparation</b>		
	Home made	93
	Ready to use not pre-poured	137
	Ready to use, plate, film, card	57
<b>1<sup>st</sup> dilution retained</b>		
	- 1	182
	- 2	95
	- 3	1
	1/40	2
	1/400	5
<b>Incubation temperature</b>		
	37±1°C	183
	30°C	95
	35°C	10
<b>Incubation duration</b>		
	20-26 h	281
	48 h	5
	30 h	1
	37 h	1
<b>Confirmatory test</b>		
	Yes	69
	No	215

<sup>(1)</sup> Similar method to NF V08-054 according to ONSSA (Office National de Sécurité Sanitaire des produits Alimentaires).

## 2.7. TOTAL COLIFORMS

236 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
NF V08-050		119
→ NM 08.0.142 <sup>(2)</sup>		8
ISO/NF ISO 4832		59
NM ISO 4832		22
AFNOR 3M		18
AFNOR BRD-07/08-12/04		4
AFNOR BIO-12/17-12/05		2
Internal method		2
Other		2
<b>Culture medium</b>		
VRBL		207
Petrifilms		19
Rapid Ecoli		5
Tempo TC		2
Other		2
<b>Preparation</b>		
Home made		90
Ready to use not pre-poured		124
Ready to use, plate, film, card		21
<b>1<sup>st</sup> dilution retained</b>		
-1		179
-2		54
-3		1
1/400		1
<b>Incubation temperature</b>		
30°C		220
37°C		16
<b>Incubation duration</b>		
20-26 h		232
48 h		3
30 h		1

AFNOR 3M method including :

3 laboratories specified utilization of AFNOR 3M-01/02-09/89 A method.

<sup>(2)</sup> Similar method to NF V 08-050 according to ONSSA.

## 2.8. THERMOTOLERANT COLIFORMS

217 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	NF V08-060	145
	→ NM 08.0.124 <sup>(3)</sup>	27
	AFNOR 3M	23
	ISO/NF ISO 4832	18
	Internal method	2
	Other	2
<b>Culture medium</b>		
	VRBL	190
	Petrifilms	24
	Other	3
<b>Preparation</b>		
	Home made	87
	Ready to use not pre-poured	106
	Ready to use, plate, film, card	23
<b>1<sup>st</sup> dilution retained</b>		
	-1	173
	-2	43
<b>Incubation temperature</b>		
	42-45°C	214
	37°C	2
	30°C	1
<b>Incubation duration</b>		
	21-25 h	214
	48 h	3

AFNOR 3M method including :

- 4 laboratories specified utilization of AFNOR 3M-01/02-09/89 C method.
- 1 laboratory specified utilization of AFNOR 3M-01/05-03/97 B method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm high sensitivity method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm EC method.

<sup>(3)</sup> Similar method to NF V08-060 according to ONSSA.

## 2.9. ESCHERICHIA COLI

308 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	ISO/NF ISO 16649-2	182
	AFNOR 3M	47
	NM ISO 16649-2	18
	AFNOR BRD-07/01-07/93	17
	AFNOR BIO-12/13-02/05	10
	AFNOR AES-10/06-01/08	9
	NM 08.0.108	8
	AFNOR BIO-12/05-01/99	4
	ISO/NF EN ISO 16649-3	3
	Internal method	3
	Other	7
<b>Culture medium</b>		
	TBX	210
	Petrifilms	48
	Rapid E. coli	21
	Rebecca	11
	Tempo EC	10
	Coli ID	6
	Other	2
<b>Preparation</b>		
	Home made	89
	Ready to use not pre-poured	163
	Ready to use, plate, film, card	56
<b>Plating method</b>		
	Surface	46
	Pour	246
	Culture medium for card	12
<b>1<sup>st</sup> dilution retained</b>		
	-1	262
	-2	36
	1/40	2
	1/400	6
<b>Incubation temperature</b>		
	41-45°C	276
	37±1°C	31
	30°C	1
<b>Incubation duration</b>		
	18-25 h	304
	48 h	4

AFNOR 3M method including :

16 laboratories specified utilization of AFNOR 3M-01/08-06/01 (*SELECT'E. COLI*) method.

1 laboratory specified utilization of AFNOR 3M-01/06-09/97 method.

2 laboratories specified utilization of AFNOR 3M Petrifilm EC method.

## 2.10. ANAEROBIC SULFITE-REDUCING BACTERIA

244 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	NF V08-061	155
	→ NM 08.0.125 <sup>(4)</sup>	15
	ISO/NF ISO 15213	47
	NM ISO 15213	14
	Internal method	7
	Other	5
<b>Culture medium</b>		
	TSC	230
	TSN	6
	Iron Sulfite agar	6
	Other	2
<b>Preparation</b>		
	Home made	90
	Ready to use not pre-poured	125
	Ready to use, plate, film, card	29
<b>Seeding way</b>		
	Plates	164
	Tubes	80
<b>1<sup>st</sup> dilution retained</b>		
	-1	220
	-2	21
<b>Incubation temperature</b>		
	44-48°C	174
	37°C	70
<b>Incubation duration</b>		
	18-24 h	198
	44-48 h	40
	72 h	6

<sup>(4)</sup> Similar method to NF V08-061 according to ONSSA.

## 2.11. CLOSTRIDIUM PERFRINGENS

188 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
ISO/NF EN ISO 7937		149
NM ISO 7937		29
NM 08.0.111		2
Internal method		1
Other		6
<b>Culture medium</b>		
TSC		187
Other		1
<b>Preparation</b>		
Home made		64
Ready to use not pre-poured		119
Ready to use, plate, film, card		5
<b>1<sup>st</sup> dilution retained</b>		
-1		180
-2		6
<b>Incubation temperature</b>		
37°C		180
44-46°C		7
34°C		1
<b>Incubation duration</b>		
18-24 h		181
48 h		7
<b>Confirmation test</b>		
None		29
Lactose-sulfite		137
Strip		9
MALDI-TOF mass spectrometry		3
Other		1

## 2.12. COAGULASE POSITIVE STAPHYLOCOCCI

303 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
<b>Method</b>		
ISO/NF EN ISO 6888-2	135	
ISO/NF EN ISO 6888-1	74	
AFNOR BKR-23/10-12/15	23	
NM ISO 6888-1	22	
AFNOR 3M-01/9-04/03	15	
AFNOR BIO-12/28-04/10	9	
NM ISO 6888-2	5	
Internal method	4	
NM 08.0.112	3	
NordVal No :049	3	
ISO/NF EN ISO 6888-3	3	
Other	5	
<b>Culture medium</b>		
RPF	131	
BP+egg yolk tellurite	94	
Easy Staph	28	
Petrifilm	16	
BP+egg yolk tellurite+ sulfamethazine	16	
Tempo STA	9	
Rapid Staph	5	
Other	2	
<b>Preparation</b>		
Home made	69	
Ready to use not pre-poured	129	
Ready to use, plate, film, cards	103	
<b>Plating method</b>		
Surface	155	
Pour	135	
Culture medium for card	10	
<b>1<sup>st</sup> dilution retained</b>		
-1	105	
-2	182	
-3	6	
1/40	4	
1/400	2	
<b>Incubation temperature</b>		
37±1°C	298	
27-30°C	4	
44°C	1	
<b>Incubation duration</b>		
40-49 h	208	
18-25 h	95	
<b>Confirmation test</b>		
None	185	
Staphylo-coagulase	81	
Clumping factor	16	
DNase	10	
MALDI-TOF mass spectrometry	3	
Other	2	

## 2.13. LISTERIA MONOCYTOGENES – ENUMERATION

249 laboratories performed the enumeration.

### RESUSCITATION

92 laboratories announce the realization of a resuscitation step.

Details concerning temperature and average duration of this step are not required anymore in the input form.

Parameters	Mode	Nb laboratories
<b>Method</b>		
	ISO/NF EN ISO 11290-2	74
	AFNOR AES-10/05-09/06	59
	AFNOR BKR-23/05-12/07	53
	NM ISO 11290-2	26
	AFNOR BRD-07/05-09/01	25
	AFNOR BRD-07/17-01/09	10
	Other	2
<b>Resuscitation medium</b>	Buffered Peptone Water or equivalent	216
	Half-fraser	26
	Fraser base	3
	Other	6
<b>Enumeration medium</b>	ALOA Count	116
	Compass Listeria	84
	Rapid Lmono	26
	AL Agar	18
	OCLA	3
	Palcam	1
	Other	1
<b>Preparation</b>	Home made	38
	Ready to use not pre-poured	51
	Ready to use, plate, film, card	159
<b>Plating method</b>	Surface	207
	Pour	39

Parameters	Mode	Nb laboratories
<b>1<sup>st</sup> dilution retained</b>	-1	158
	-2	88
<b>Incubation temperature</b>	37°C	243
	30°C	5
	20°C	1
<b>Incubation duration</b>	44-49 h	207
	24-27 h	41
	15 h	1
<b>Confirmation test</b>	None	47
	Biochemical	148
	Biochemical + CAMP	37
	MALDI-TOF mass spectrometry	3
	Other	10
<b>Nb of colonies tested per plate</b>	1	63
	2-4	14
	5	111
	6	2

## 2.14. SALMONELLA – DETECTION

**312** laboratories performed the detection.

Methods used by laboratories are clarified in the following table :

Parameter	Mode	Nb laboratories
<b>Method</b>		
ISO/NF EN ISO 6579-1		81
AFNOR BKR 23/07-10/11 (IRIS Salmonella)		77
NM ISO 6579-1		36
AFNOR BRD 07/11-12/05 (Rapid Salmonella)		31
AFNOR BIO 12/32-10/11 (VIDAS SPT)		24
AFNOR BIO 12/41-03/17 (SALMA One day)		24
AFNOR BIO 12/16-09/05 (VIDAS Easy Salmonella)		16
AFNOR BIO 12/01-04/94 (VIDAS SLM)		7
AFNOR UNI 03/07-11/13 (PCR)		3
AFNOR BRD 07/06-07/04 (PCR)		2
AFNOR BIO 12/38-06/16 (GENE UP Salmonella)		2
AFNOR TRA 02/08-03/01 (TRANSIA PLATE Salmonella GOLD)		2
Internal method		1
Other		6

No detail of methodology was asked to laboratories using other methods than ISO/NF EN ISO 6579-1 and NM ISO 6579-1 methods and proposed in the questionnaire.

You will find below a short description of these methods :

Method	Pre-enrichment	Enrichment	Isolation
AFNOR BKR 23/07-10/11 <b>IRIS Salmonella</b>		IRIS Salmonella Enrichment / 41,5°C - 18±2h	IRIS / 37°C - 24±3h
AFNOR BRD 07/11-12/05 <b>Rapid Salmonella</b>		BPW + Salmonella capsule / 41,5°C - 18±2h	Rapid Salmonella / 37°C - 24±2h
AFNOR BIO 12/32-10/11 <b>VIDAS SPT</b>		BPW + Salmonella supplement / 41,5°C - 18/24h	Chrom ID / 37°C - 24h
AFNOR BIO 12/41-03/17 <b>SALMA One day</b>		BPW + Salmonella supplement / 41.5°C – 16/24h	SALMA / 37°C - 24±3h
AFNOR BIO 12/16-09/05 <b>VIDAS Easy Salmonella</b>	BPW / 37°C - 16/20h	SX2 / 41,5°C - 22/26h	Chrom ID / 37°C - 24h
AFNOR BIO 12/01-04/94 <b>VIDAS SLM</b>	BPW / 35°C – 24±2h	Tetrathionate (42°C - 6/8h) – Selenite cystine (35-37°C – 6/8h) + M-Broth (42°C – 18h)	Vidas Heat & Go
AFNOR UNI 03/07-11/13 <b>PCR</b>		BPW + supplement / 34-38°C – 20/24h	Lysis + PCR
AFNOR BRD 07/06-07/04 <b>PCR</b>		BPW / 37°C – 18/21h	Lysis + PCR
AFNOR BIO 12/38-06/16 <b>GENE UP Salmonella</b>		BPW / 42°C – 18/24h	Lysis + PCR
AFNOR TRA 02/08-03/01 <b>TRANSIA PLATE Salmonella GOLD</b>	BPW / 37°C – 16/20h	RVS / 41.5°C – 18/24h	ELISA test

The detail of the methodology followed by 117 laboratories using ISO/NF EN ISO 6579-1 and NM ISO 6579-1 method, and the 7 laboratories using an internal or another method, is clarified in the following table :

Parameter	Mode	Nb laboratories
<b>Method</b>		
	ISO/NF EN ISO 6579-1	81
	NM ISO 6579-1	36
	Internal method	1
	Other	6
<b>Pre-enrichment medium</b>		
	None pre-enrichment	2
	Buffered Peptone Water	118
	Other	4
<b>Pre-enrichment temperature</b>		
	37±1°C	115
	42-42.5°C	4
	22°C	2
	30°C	1
<b>Pre-enrichment duration</b>		
	16-20 h	90
	22-24 h	32
<b>Enrichment medium</b>		
	None enrichment	5
	RVS	113
	MKTn	108
	Selenite-cystine broth	26
	Other	3
<b>Isolation medium</b>		
	XLD	107
	Hektoen	30
	Bismuth Sulfate	25
	IRIS Salmonella agar	15
	ASAP	14
	GVB	12
	SS	10
	Brilliance Salmonella	7
	Rapid Salmonella	5
	Compass Salmonella	5
	Rambach	2
	Other	12
<b>Confirmation test</b>		
	Biochemical	46
	Biochemical + serological agglutination	67
	MALDI-TOF mass spectrometry	5
	Other	4

## 2.15. LISTERIA MONOCYTOGENES – DETECTION

283 laboratories performed the detection.

Parameter	Mode	Nb laboratories
<b>Method</b>		
AFNOR AES 10/03-09/00 (ALOA one day)		57
ISO/NF EN ISO 11290-1		57
AFNOR BKR 23/02-11/02 (Compass L. mono)		57
NM ISO 11290-1		31
AFNOR BRD 07/04-09/98 (Rapid' L. mono)		23
AFNOR BRD 07/16-01/09 (Agar Listeria)		11
AFNOR BIO 12/11-03/04 (VIDAS LMO2-37°C)		9
AFNOR BIO 12/27-02/10 (VIDAS LMX)		8
AFNOR BIO 12/02-06/94 (VIDAS Listeria)		7
AFNOR BIO 12/18-03/06 (VIDAS LDUO)		4
AFNOR BRD 07/10-04/05 (IQ Check Listeria)		4
AFNOR UNI 03/08-11/13 (PCR)		3
AFNOR BIO 12/40-11/16 (GENE UP LMO)		3
AFNOR UNI 03/04-04/05 (Listeria Precis)		3
Internal method		2
Other		4

No detail of methodology was asked to laboratories using other method than ISO/NF EN ISO 11290-1 and NM ISO 11290-1 methods and proposed in the questionnaire.

You will find below a short description of these methods :

Méthod	Primary enrichment		Secondary enrichment		Isolation
	Medium	Incubation	Medium	Incubation	
AFNOR AES 10/03-09/00 <b>ALOA one day</b>	Half-Fraser	30°C - 24±2h			ALOA One Day 37°C – 24/48h
AFNOR BKR 23/02-11/02 <b>Compass L. mono</b>	Half-Fraser	30°C - 24±2h			Compass Listeria Agar 37°C – 24h
AFNOR BRD 07/04-09/98 <b>Rapid' L. mono</b>	Half-Fraser	30°C - 24±2h			Rapid L'mono 37°C – 24h
AFNOR BRD 07/16-01/09 <b>Agar Listeria</b>	Half-Fraser	30°C - 24±2h			Agar Listeria 37°C – 24h
AFNOR BIO 12/11-03/04 <b>VIDAS LMO2 (37°C)</b>	Half-Fraser	30°C - 24/26h	Fraser	37°C - 24/26h	Chromogenic medium / Palcam / Oxford
AFNOR BIO 12/27-02/10 <b>VIDAS LMX</b>	LMX	37°C - 26/30h			ChromID 37°C – 24h
AFNOR BIO 12/02-06/94 <b>VIDAS Listeria</b>	Half-Fraser	37°C - 26/30h	Fraser	30°C - 24/26h	Palcam et Oxford 37°C – 24h
AFNOR BIO 12/18-03/06 <b>VIDAS LDUO</b>	LX	30°C - 24±2h	LX	30°C - 24/26h	Chromogenic medium / Palcam / Oxford
AFNOR BRD 07/10-04/05 <b>IQ Check Listeria</b>	Half-Fraser / LSB	30°C – 23/25h			Lysis + PCR
AFNOR UNI 03/08-11/13 <b>PCR</b>	LEB	37°C - 24/28h			Lysis + PCR
AFNOR BIO 12/40-11/16 <b>GENE UP LMO</b>	LPT	35-37°C - 24±2h			ALOA 35-37°C – 24/48h
AFNOR UNI 03/04-04/05 <b>Listeria Precis</b>	One Broth Listeria	30°C - 24±2h			Brilliance Listeria 37°C – 24h

The detail of the methodology followed by 88 laboratories using NF ISO/EN ISO 11290-1 and NM ISO 11290-1 methods, and the 6 laboratories using an internal or another method, is clarified in the following table :

Parameter	Mode	Nb laboratories
<b>Method</b>		
ISO/NF EN ISO 11290-1	57	
NM ISO 11290-1	31	
Internal method	2	
Other	4	
<b>Primary enrichment medium</b>		
None primary enrichment	0	
Half-Fraser	86	
One broth Listeria	1	
Other	7	
<b>Primary enrichment temperature</b>		
30°C	86	
37°C	6	
22°C	1	
<b>Primary enrichment duration</b>		
21-27 h	91	
28 h	1	
48 h	1	
<b>Secondary enrichment medium</b>		
None secondary enrichment	7	
Fraser	86	
Other	1	
<b>Secondary enrichment temperature</b>		
37°C	83	
30°C	2	
22°C	1	
<b>Secondary enrichment duration</b>		
23-27 h	72	
48 h	14	
<b>Isolation medium</b>		
Palcam	67	
Ottaviani et Agosti	52	
Compass Listeria	37	
Oxford	10	
Rapid L'mono	7	
Brilliance Listeria	1	
Other	0	
<b>Isolation temperature</b>		
37°C	90	
30°C	2	
<b>Isolation duration</b>		
44-48 h	54	
22-24 h	38	
<b>Confirmation test</b>		
None	3	
Biochemical	62	
Biochemical + CAMP	22	
MALDI-TOF mass spectrometry	3	
Other	3	
<b>Nb of colonies per plate</b>		
1	36	
2-3	7	
5	41	

### 3. ASSESSMENT OF PERFORMANCE (INDIVIDUEL REPORTS)

#### 3.1. PERFORMANCES IN ENUMERATION

Performance is assessed on two criteria : **precision and trueness**.

The assigned value of the contamination used to assess the trueness and the reference standard deviation for the assessment of the precision are consensual values obtained with the results of all the participants. These values are obtained by robust estimation methods in order to eliminate influence of aberrant results. However some results are excluded of the statistical analysis. That is the case when laboratories do not give results for all contaminated units, when results are "less than CFU/g", when samples are analyzed after the deadline (time of receipt > 4 days after sending or time of analysis >15 days after sending) or when this information is not specified.

A statistical analysis has also be done to highlight potential relations between techniques used (delay of analysis, preservation temperature, preparation of the initial suspension, homogenization technique, resuscitation conditions, method used, media used, manufacturers of media, preparation mode, plating method, incubation conditions) and results obtained. We need to clarify that this statistical link is not involved in a cause - effect relationship. Indeed, this link may be due to a not documented factor.

When a significant statistical link is identified between use of a technique and the obtained results, the assessment of performance is done considering the influence of one or several factors involved if their effect translates into a contamination's difference higher than 0.15 log CFU/g for non-selective media or higher than 0.30 log CFU/g for selective media (these limits match with productivity limits of culture media usually recommended in the standard NF EN ISO 11133).

#### PRECISION

The precision reflects the repeatability (or reproducibility intra-laboratory) of your work.

The standard deviation of your results,  $s$ , is compared to the robust estimation of the standard deviation (reference standard deviation of precision),  $s^*$ , obtained with algorithm S from the standard ISO 13528 applied to all standard deviations obtained by laboratories included in the statistical analysis.

An index score is then calculated using the following formula :  $i = (k-1) \cdot \frac{s^2}{s^{*2}}$  (with  $k$ , number of contaminated units and retained in the statistical analysis, usually 5 ).

The standard ISO 13528 do not provide warning and action limits for this score, so its interpretation is left to your discretion.

As an indicator, we suggest following values by analogy with those indicated for the evaluation of trueness.

For  $k=5$ , a score lower than 0.1 or higher than 18 may be considered as an action signal and a score lower than 0.45 or higher than 11.5 may be considered as a warning signal.

For  $k=4$ , a score lower than 0.03 or higher than 15.5 may be considered as an action signal and a score lower than 0.2 or higher than 9.5 may be considered as a warning signal.

For  $k=3$ , a score lower than 0.003 or higher than 13.2 may be considered as an action signal and a score lower than 0.05 or higher than 7.5 may be considered as a warning signal.

For  $k=2$ , a score lower than 0.000002 or higher than 10.3 may be considered as an action signal and a score lower than 0.0008 or higher than 5.2 may be considered as a warning signal.

## TRUENESS

The trueness reflects the closeness of the mean of your results to the contamination's assigned value of samples. It has been evaluated for all enumerated flora.

The mean of your results in log CFU/g,  $m$  (on contaminated units and included in the statistical analysis), is compared to the contamination's assigned value,  $m_{pt}$ , obtained with algorithm A from the standard ISO 13528 applied to all laboratories mean included in the statistical analysis. When groups are formed, each one is characterized by its own assigned value.

A z score is then calculated with the following formula :  $z = \frac{m - m_{pt}}{\sigma_{pt}}$ , where  $\sigma_{pt}$  is the standard deviation for proficiency assessment (robust estimation of the standard deviation obtained by participants).

The standard ISO 13528 specifies that a z score lower than -3 or higher than +3 must be considered as an action signal and that a z score lower than -2 or higher than +2 must be considered as a warning signal.

In this report, we specify, estimations of interlaboratories standard deviation for enumerations proposed as well as reproducibility standard deviation or global standard deviation for the test (parameters including interlaboratories variability and the variability of the precision).

## INDIVIDUAL REPORTS – FOR EACH CRITERIA YOU FIND THE FOLLOWING INFORMATIONS

- your results in logarithm base 10 (-1 when the answer is < limit and NaN when there is no answer).  
Comment : the presentation order of your results does not necessarily correspond to the order you sent them, this order is the same for all the flora.
- histogram for the studied parameter (laboratories standard deviations for the precision and laboratory's means for the trueness) with an asterisk indicating the location of your result,
- standard deviation (precision) or mean (trueness) of your results (on contaminated units and retained in the statistical analysis),
- the method declared in your results input,
- when necessary, your group in relation to the technique used,
- precision score or z score,
- number of laboratories which made analysis (and belonging to your group),
- number of laboratories included in the statistical analysis,
- reference standard deviation for the precision or assigned value of the contamination and standard deviation aptitude assessment (trueness),
- number of laboratories with a satisfactory signal,
- number of laboratories with a warning signal,
- number of laboratories with an action signal.

### 3.1.1. MICROORGANISMS AT 30°C

None significant effect of the analysis technique has been highlighted.

Microorganisms at 30°C	
Assigned value of the contamination (log cfu/g)	5.469
Assigned value uncertainty (log cfu/g)	0.0060
Standard deviation for proficiency assessment (log cfu/g)	0.0843
Standard deviation for precision (log cfu/g)	0.0530
Interlaboratory's standard deviation (log cfu/g)	0.0809
Reproducibility standard deviation (log cfu/g)	0.0967

### 3.1.2. ENTEROBACTERIACEAE

A significant “effect” of the culture medium, manufacturer and the retained dilution has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in three groups :

Enterobacteriaceae	Group 1	Group 2	Group 3
Assigned value of the contamination (log cfu/g)	3.003	3.241	3.429
Assigned value uncertainty (log cfu/g)	0.0186	0.0364	0.0229
Standard deviation for proficiency assessment (log cfu/g)	0.1948	0.2039	0.1307
Standard deviation for precision (log cfu/g)		0.0856	
Interlaboratory's standard deviation (log cfu/g)	0.1910	0.2002	0.1249
Reproducibility standard deviation (log cfu/g)	0.2093	0.2178	0.1514

### 3.1.3. TOTAL COLIFORMS

A significant “effect” of the culture medium, manufacturer, and the retained dilution has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in three groups :

Total coliforms	Group 1	Group 2	Group 3
Assigned value of the contamination (log cfu/g)	2.915	3.078	3.366
Assigned value uncertainty (log cfu/g)	0.0211	0.0334	0.0410
Standard deviation for proficiency assessment (log cfu/g)	0.1999	0.1907	0.1672
Standard deviation for precision (log cfu/g)		0.0775	
Interlaboratory's standard deviation (log cfu/g)	0.1969	0.1875	0.1636
Reproducibility standard deviation (log cfu/g)	0.2147	0.2061	0.1846

### 3.1.4. THERMOTOLERANT COLIFORMS

A significant “effect” of the retained dilution has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in two groups :

Thermotolerant coliforms	Group 1	Group 2
Assigned value of the contamination (log cfu/g)	2.922	3.288
Assigned value uncertainty (log cfu/g)	0.0177	0.0369
Standard deviation for proficiency assessment (log cfu/g)	0.1782	0.1890
Standard deviation for precision (log cfu/g)	0.0884	
Interlaboratory's standard deviation (log cfu/g)	0.1737	0.1848
Reproducibility standard deviation (log cfu/g)	0.1937	0.2037

### 3.1.5. *ESCHERICHIA COLI*

A significant “effect” of the preparation mode of the culture medium and the retained dilution has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

<i>Escherichia coli</i>	
Assigned value of the contamination (log cfu/g)	2.899
Assigned value uncertainty (log cfu/g)	0.0150
Standard deviation for proficiency assessment (log cfu/g)	0.2046
Standard deviation for precision (log cfu/g)	0.0877
Interlaboratory's standard deviation (log cfu/g)	0.2008
Reproducibility standard deviation (log cfu/g)	0.2192

### 3.1.6. ANAEROBIC SULFITE-REDUCING BACTERIA

Only units n°3 and 4 were artificially contaminated.

None significant effect of the analysis technique has been highlighted.

Anaerobic sulfite-reducing bacteria	
Assigned value of the contamination (log cfu/g)	2.112
Assigned value uncertainty (log cfu/g)	0.0171
Standard deviation for proficiency assessment (log cfu/g)	0.2022
Standard deviation for precision (log cfu/g)	0.1078
Interlaboratory's standard deviation (log cfu/g)	0.1873
Reproducibility standard deviation (log cfu/g)	0.2161

Comment :

- 5 laboratories detected ASR in unit n°1 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 3800 cfu/g.
- 8 laboratories detected ASR in unit n°2 non-artificially contaminated by *C. perfringens* with a contamination level from 1 to 4300 cfu/g.
- 13 laboratories detected ASR in unit n°5 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 5200 cfu/g.

### 3.1.7. *CLOSTRIDIUM PERFRINGENS*

Only units n°3 and 4 were artificially contaminated.

None significant effect of the analysis technique has been highlighted.

<b><i>Clostridium perfringens</i></b>	
Assigned value of the contamination (log cfu/g)	2.105
Assigned value uncertainty (log cfu/g)	0.0158
Standard deviation for proficiency assessment (log cfu/g)	0.1660
Standard deviation for precision (log cfu/g)	0.0995
Interlaboratory's standard deviation (log cfu/g)	0.1504
Reproducibility standard deviation (log cfu/g)	0.1803

Comment :

- 3 laboratories detected *C. perfringens* in unit n°1 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 2400 cfu/g.
- 5 laboratories detected *C. perfringens* in unit n°2 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 4400 cfu/g.
- 5 laboratories detected *C. perfringens* in unit n°5 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 4000 cfu/g.

### 3.1.8. COAGULASE POSITIVE STAPHYLOCOCCI

A significant “effect” of the resuscitation’s duration has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

<b>Coagulase positive Staphylococci</b>	
Assigned value of the contamination (log cfu/g)	3.836
Assigned value uncertainty (log cfu/g)	0.0124
Standard deviation for proficiency assessment (log cfu/g)	0.1669
Standard deviation for precision (log cfu/g)	0.0657
Interlaboratory's standard deviation (log cfu/g)	0.1643
Reproducibility standard deviation (log cfu/g)	0.1769

### 3.1.9. *LISTERIA MONOCYTOGENES*

Only units n°3, 4 and 5 were artificially contaminated.

A significant “effect” of the culture medium has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

<b><i>Listeria monocytogenes</i></b>	
Assigned value of the contamination (log cfu/g)	3.514
Assigned value uncertainty (log cfu/g)	0.0083
Standard deviation for proficiency assessment (log cfu/g)	0.1019
Standard deviation for precision (log cfu/g)	0.0724
Interlaboratory's standard deviation (log cfu/g)	0.0929
Reproducibility standard deviation (log cfu/g)	0.1178

### 3.2. PERFORMANCES IN DETECTION

The performance is assessed by the capacity to detect only samples contaminated by *Salmonella* and *Listeria monocytogenes* (no false positive or false negative results).

#### 3.2.1. DETECTION – *SALMONELLA*

Only units n°3, 4 and 5 were artificially contaminated.

297 laboratories obtained correct results.

8 laboratories obtained false positive results (respectively 5 and 5 false-positive for units n° 1 and 2).

13 laboratories obtained false negative results (respectively 4, 3 and 8 false-negative for units n° 3, 4 and 5).

#### 3.2.2. DETECTION – *LISTERIA MONOCYTOGENES*

Only units n°3, 4 and 5 were artificially contaminated.

274 laboratories obtained correct results.

6 laboratories obtained false positive results (respectively 2 and 4 false-positive for units n°1 and 2).

4 laboratories obtained false negative results (respectively 1, 0 and 3 false-negative for units n°3, 4 and 5).

### 3.3. EVOLUTION OF PERFORMANCE

You will find, on each page of your performance's assessment, a graph representing evolution of it on different tests since the 54<sup>th</sup> scheme.

In order to interpret your control card with z scores, you can refer to the standard ISO 13528 §10.8.2.2, explaining the 3 « out of control » situations:

- Just one overtaking of the action limit ( $z < -3$  or  $z > 3$ ),
- 2 consecutives z scores out of 3 overtaking of the warning limit ( $2 < z < 3$  or  $-3 < z < -2$ ),
- 6 consecutives z scores regularly increasing or decreasing.